

OK
CMA 225675

Service Repair Sheet SRS67548

Contact Name
Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Seri Cooke

Emergency Medical Retrieval Transfer Service (EMRTS)

EMRTS Support Officer

0300 3000 057

00000159

seri.cooke@wales.nhs.uk

03/Aug/2020

Catherine Spence

Viamed

Quote

O/N
92065767
Cg 9.9.2020

Notes 29/Jul/2020 Kate Griffiths

29/Jul/2020 Kate Griffiths

S/n: A1708160003

29/Jul/2020 Kate Griffiths

S/n: A1708160003 "On power up device failed. Batteries changed, same fault. No evidence of physical damage. On return to base powered up but with red screen. SPO2 connection trialled, HR reading but no SATS. Customer request: Could the following ID be included on any communications please: 01613

03/Aug/2020 Catherine Spence

unit only with silicone cover

Ready For quote

CGNEEN 27.8.2020

Repair Complete Signed

CGNEEN 9.9.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32963	Capnograph	4410520	A1708160003	N

4440000 x @ £90 0.25hr = £45
S/N, SRS, SRN

4400100 x 1 @ ~~£290~~ 290
SRS, SRN

UPS x 1 @ £10