

Miniox

Neil Bell's

11/9/00


2 prototypes

eg) battery duration test.

Software for +ve + -ve alarms.

Click of piezo - to be removed. - capacitor

~~Link to be possibly~~ to be incorporated
~~added later.~~



Reset switch may be incorporated
if necessary.

Spike caused by display may cause problems with
alarm or duration due to increased time when
proc awake and waiting to sample.

Batt alarm @ 2.4V. ~~is~~ generated by wake-up IC.

New displays (standard software) to be ordered.

Display needs setting & cal pot to be user changed.

Alm 7²65. $\Rightarrow$ standard. (2). $\times 1$.
one for Neil (Laser).

—— u ——

Michael to mount

Sort out case. —— (metal).

the ϕ -ve alarms to be sorted.

—— 1 ——

Batt duration test 11/9/00 10-10Am.
(26mSH).

25 days? Test alarm carried out
 $\hookrightarrow$ to be c/o every Monday morn.

prefer louder pings.