

Service Repair Sheet SRS67569

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Sumera Ahmed

East Surrey Hospital

01737 768511 ext 6600

00004300

sumera.ahmed@nhs.net

09/Sep/2020

Robert Connor

Viamed

Warranty

Notes 20/Aug/2020 Sarah Walton

20/Aug/2020 Sarah Walton

Serial number: PR02094A26

Returning their V1000 after it s service. Customer says - I inserted a fresh set of batteries and the heart-shaped button doesnâ€™t work and it wont turn off unless the batteries are taken out. I took those batteries out and inserted another fres set and now it wonâ€™t turn on at all.

09/Sep/2020 Robert Connor

1 x V1000 s/n PR02094A26

Ready For quote

Repair Complete Signed

C Connor 30.9.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33110	Foetal Simulator	1410000	PR02094A26	Yes (r)

1480000 x 1 fOC
S/N, SRS, SRN

1430295 x 1 fOC
SRS, SRN

UPS x 1 fOC