
Fwd: Replacing Tom Thumb valves with Blenders in Viamed cabinets in Maternity Dept

1 message

Main Account <office@viamed.co.uk>
To: Zoey Teal <zoey.teal@viamed.co.uk>

Wed, Sep 30, 2020 at 2:48 PM

----- Forwarded message -----

From: **Steve Hardaker** <office@viamed.co.uk>
Date: Wed, 30 Sep 2020 at 14:44
Subject: Fwd: Replacing Tom Thumb valves with Blenders in Viamed cabinets in Maternity Dept
To: Main Account <office@viamed.co.uk>

Hi,

Please can someone send a quotation to John Wall at South Tyneside General for the following items, his full details are in the email below:

- 0310200 - Air/oxygen blender kit (inc. hoses, rail mount and flowmeter) @ £1,102.50+VAT
- 0332057 - Rail upgrade kit, 40cm (includes fixings) @ £40.00+VAT
- 0310250 - Tom Thumb conversion kit (parts only, for self-modification) @ £21.78+VAT
- 0330035 - Clear tubing 1m @ £1.50+VAT

In the email below, I recommended 2x Rail upgrade kits per cabinet but have since advised him that we can get away with 1 (we are running out of rails and I'm not sure we can get more).

Please email it on to him and cc me in, thanks. Skype me if any queries as I have already set an out of office until Tuesday to do service visits.

Thanks,

Steve H

----- Forwarded message -----

From: **Wall John** <John.Wall@stft.nhs.uk>
Date: Tue, 29 Sep 2020 at 09:21
Subject: RE: Replacing Tom Thumb valves with Blenders in Viamed cabinets in Maternity Dept
To: steve.hardaker@viamed.co.uk <steve.hardaker@viamed.co.uk>

Hi Steve,

Thanks for your reply, I'd like to order a kit to upgrade one of the cabinets initially then I'll need to submit a business case to get the funding for the other 8 cabinets.

Could you send the Tom Thumb service manuals so we can do the conversion here.

Can you supply a quotation for 1 cabinet upgrade (Air/oxygen blender kit, Rail upgrade kit) plus Tom Thumb conversion kit?

Many thanks

John

Senior Clinical Engineer

South Tyneside General Hospital

Harton Lane

South Shields

NE34 0PL

01914041104

Hi John,

We can supply a rail mounted air/oxygen blender that will fit into the infant resuscitation cabinet, but in order to connect the blender to the Tom Thumb, the Tom Thumb must be modified to replace the oxygen flowmeter and hose with a tapered low-pressure gas inlet.

As your systems were provided in 2003, the medirails will also need replacing with longer sections in order to accommodate the blender. These are available as a kit with the fixings.

I have attached a document that details what is required and the options for achieving it, although, please note, this is an older document and we are no longer providing on-site service visits. The items are currently priced as follows:

- 0310200 - Air/oxygen blender kit (inc. hoses, rail mount and flowmeter) @ £1,102.50+VAT
- 0332057 - Rail upgrade kit, 40cm @ £40.00+VAT (2 required per cabinet, includes fixings)
- 0310250 - Tom Thumb conversion kit @ £21.78+VAT (parts only, for self-modification)

If you want to return the Tom Thumbs to us to modify, this will require an additional service cost per unit as follows:

- 0380000 - Tom Thumb service @ £90.00+VAT
- Return carriage: £10.00+VAT

Please have a read through of the document and let me know if you have any queries. If you want a quotation, please just let me know how many systems need the upgrade and whether you would like to return the Tom Thumb to us or do the modification yourselves.

If you would like me to send a service manuals for the Tom Thumbs so that you know what you are dealing with, please let me know.

Regards,

Steve Hardaker
Technical Support Manager
Viamed Ltd.

Please note: Viamed is now enacting a coronavirus contingency plan to allow sales and admin Staff to work from home, and I am now working remotely. Telephone calls to the main office will not be answered, but emails will. Orders may take longer to process but we aim to continue to operate and serve our customers as best we can.

Website: www.viamed.co.uk
email: steve.hardaker@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up
Twitter: twitter.com/ViamedLtd
Facebook: facebook.com/ViamedLtdUK
Company registered in England, No. 1291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542
GDPR / Privacy Policy available at www.viamed.co.uk/policy

Hi,

I was hoping that you could help. We need to replace the Tom Thumb valves in 9 Maternity resus cabinets with Oxygen blenders and I need one fitting to an infant resuscitaire in Paeds resus dept.

I'm not sure what is currently available and what would physically fit in our cabinets.

Is there anyone available that could advise me please?

Kind regards

John Wall
Senior Clinical Engineer
South Tyneside General Hospital
Harton Lane
South Shields
NE34 0PL

01914041104

