

Service Repair Sheet 67575

Contact Name	Michelle Anderson
Company/ Hospital Name	V-Tech UK Ltd
Department	Accounts Department
Position	Accounts Controller
Direct Phone	0208 498 1288
General Phone	0208 498 1288
Opera Account	00012679
Email	accounts@v-techuk.com
Date Received	01/Sep/2020
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33090	0110122	402145	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110122 Oxygen sensor - Automotive R-22A

11.8 in air and 55.2 in oxygen, the sensor is nonlinear and replaced under warranty with S/N 415987
