

Service Repair Sheet 67517

Contact Name	Maria Hewitt
Company/ Hospital Name	Eurotec Equipment Services
Position	Ms Hewitt
General Phone	01948 780546
Opera Account	00020096
Email	maria-hewitt@btconnect.com
Date Received	13/Jul/2020
Booked in By	Catherine Spence

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN32938	0110132	376943	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110132 Oxygen sensor - Automotive R-22AV

13.6 in air and 64.3 in oxygen, this sensor is high and replaced under warranty with S/N 417599

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN32939	0110132	376946	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110132 Oxygen sensor - Automotive R-22AV

13.5 in air and 64.8 in oxygen, this sensor is high and replaced under warranty with S/N 525370