

Service Repair Sheet SRS67594

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical Ltd

Department

Position

Direct Phone

01903 522911

General Phone

01903 522911

Opera Account

00011982

Email

ashley@thamesmedical.com

Order Number

PO-01368

Date Received

28/Sep/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 18/Sep/2020 Sophie Lines

18/Sep/2020 Sophie Lines

18/Sep/2020 Sophie Lines

Has two pieces of equipment for quote/repair - I have asked for serial numbers/details

18/Sep/2020 Sophie Lines

IRMA â€œ SN230482 â€œ Reported as Co2 detector not working

28/Sep/2020 Catherine Spence

received in with airway adapter

Ready For quote

CG Neen 1-10-2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33124	Accessory	4420505	230482	N

4420505 x 1
S/SRS, SRN

£807.50

UPS x 1

£10.