

Test Certificate

regarding
Electrical Safety and Electromagnetic Compatibility Requirements




CEcert GmbH, Alter Holzhafen 19, 23966 Wismar (Germany)

Test Laboratory

Certificate No.: 608.029.1

License Holder:	Bluepoint MEDICAL GmbH&Co.KG	
	An der Trave 15	
	23923 Selmsdorf, Germany	
Product identification::	SpO2-Monitor with Finger-Sensor	
Type/Model/Parameters:	OxyTrue A with SoftCap-Sensor SC6500	
P/N; S/N:	P/N: 1020112001	
	S/N: 00000152	
Test standards:	IEC 60601-1:1988+A1:1991+A2:1995	
	IEC 60601-1-2:2007	
	with applicable additional requirements of ISO 9919:2005	
Classifications and Restrictions:		
Electrical Safety	Protection against electric shock:	Internally powered equipment
	Applied part:	Typ BF
	Environmental conditions:	Ta = 40°C
	Dielectric strength:	1500 V (1min)
EMC	Emission:	Group 1 Class B
	Immunity:	not life supporting
Test reports:	408.029.1 Improvements, 408.029.2 Electrical Safety,	
	407.202.1 EMC	
Test result:	PASS	

This test certificate is relevant exclusively to the item(s) submitted for testing.
Consider the test report specified in this certificate and the safety indications specified in the technical documentation to the product.



Head of test laboratory

Wismar, Wednesday, 21 May 2008

Dipl.-Ing. Bernd Schmidt