

Service Repair Sheet SRS67487

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Tracy Wang

Company/ Hospital Name

Justec

Department

Position

Direct Phone

General Phone

075511426068930/1/2

Opera Account

Email

tracy.wang@justec.cn

Order Number

Date Received

27/Jun/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 12/May/2020 Sarah Walton

12/May/2020 Sarah Walton

Serial Number 364241

Faulty sensor

RS advised that is replacement is to be shipped with their next VN202 shipment

12/May/2020 Sarah Walton

12/May/2020 Sarah Walton

12/May/2020 Sarah Walton

Serial Number 364241

Faulty sensor

Ignore the comment about RS - I read the email wrong!

27/Jun/2020 Catherine Spence

UFO sensor s/n 364241 received

Ready For quote

Repair Complete Signed

CC. Neen 8.10.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32914	Oxygen Sensor	0110081	364241	Yes

0140013 x 1 FCC
SN, SRS, SEN

0110081 x 1 FCC
New SN, SRS, SEN.

UPS x 1 FCC.