

Service Repair Sheet SRS67495

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Tracy Wang

Company/ Hospital Name

Justec Shenzhen Co. Ltd

Department

Position

Direct Phone

General Phone

0086 755 2606 8931

Opera Account

00006027

Email

tracy.wang@justec.cn

Order Number

Date Received

27/Jun/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 22/May/2020 Ryan Swaine

22/May/2020 Ryan Swaine

22/May/2020 Ryan Swaine

The SN for the 9 pcs R-17MED:

416898, 416838, 416843, 416866, 416595, 416977, 416842, 416873, 416849

Failed with various faults including very high output & faulty connectors

02/Jun/2020 Ryan Swaine

22/May/2020 Ryan Swaine

22/May/2020 Ryan Swaine

22/May/2020 Ryan Swaine

The SN for the 9 pcs R-17MED:

416898, 416838, 416843, 416866, 416595, 416977, 416842, 416873, 416849

Failed with various faults including very high output & faulty connectors

02/06/2020 Another 9x R-17MED reported as faulty: 417183, 416791, 417189, 417002, 419893, 416613, 416976, 419947, 416712.

27/Jun/2020 Catherine Spence

received 19 items

Ready For quote

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32896	Oxygen Sensor	0110017	416898	Yes
SRN32897	Oxygen Sensor	0110017	416866	Yes
SRN32898	Oxygen Sensor	0110017	416849	Yes
SRN32899	Oxygen Sensor	0110017	416842	Yes
SRN32900	Oxygen Sensor	0110017	416873	Yes
SRN32901	Oxygen Sensor	0110017	416838	Yes
SRN32902	Oxygen Sensor	0110017	416595	Yes
SRN32903	Oxygen Sensor	0110017	416843	Yes
SRN32904	Oxygen Sensor	0110017	416977	Yes
SRN32905	Oxygen Sensor	0110017	419947	Yes
SRN32906	Oxygen Sensor	0110017	417183	Yes
SRN32907	Oxygen Sensor	0110017	416976	Yes

0140013 x 18 fcc
S/N, SRS, SEN

0110017 x 18 fcc
New S/N, SRS, SEN

UPS x 1 fcc

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32908	Oxygen Sensor	0110017	417189	Yes
SRN32909	Oxygen Sensor	0110017	419893	Yes
SRN32910	Oxygen Sensor	0110017	416712	Yes
SRN32911	Oxygen Sensor	0110017	416613	Yes
SRN32912	Oxygen Sensor	0110017	417002	Yes
SRN32913	Oxygen Sensor	0110017	416791	Yes