

Service Repair Sheet 67564
 Contact Name Tracy Wang
 Company/ Hospital Name Justec Shenzhen Co. Ltd
 Direct Phone 0755-26068930/1/2
 General Phone 0086 755 2606 8931
 Opera Account 00006027
 Email tracy.wang@justec.cn
 Date Received 01/Sep/2020
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33025	0110022	460412	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.8 in air and 41.8 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520573

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33026	0110022	443058	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

9.6 in air and 45.8 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520574

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33027	0110022	467948	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

9.0 in air and 42.9 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520575

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33028	0110022	467967	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.4 in air and 40.3 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520576

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33029	0110022	468015	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

7.8 in air and 37.3 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520577

Service Repair Sheet 67564
 Contact Name Tracy Wang
 Company/ Hospital Name Justec Shenzhen Co. Ltd
 Direct Phone 0755-26068930/1/2
 General Phone 0086 755 2606 8931
 Opera Account 00006027
 Email tracy.wang@justec.cn
 Date Received 01/Sep/2020
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33030	0110022	468048	Oxygen Sensor	
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.6 in air and 41.5 in oxygen. This sensor is nonlinear replaced under warranty with S/N 520578

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33031	0110022	468124	Oxygen Sensor	
				Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.9 in air and 42.2 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520579

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33032	0110022	468140	Oxygen Sensor	
				Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.7 in air and 41.6 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520580

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33033	0110022	468201	Oxygen Sensor	
				Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

8.7 in air and 41.2 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520581

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33034	0110022	468281	Oxygen Sensor	
				Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

9.0 in air and 42.8 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520582

Service Repair Sheet 67564
 Contact Name Tracy Wang
 Company/ Hospital Name Justec Shenzhen Co. Ltd
 Direct Phone 0755-26068930/1/2
 General Phone 0086 755 2606 8931
 Opera Account 00006027
 Email tracy.wang@justec.cn
 Date Received 01/Sep/2020
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33035	0110022	473632	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

7.8 in air and 37.2 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 520583

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33036	0110017	466052	Oxygen Sensor	
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Connector fault. Replaced under warranty with S/N 510048

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33037	0110017	469717	Oxygen Sensor	
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Connector fault. Replaced under warranty with S/N 522589

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33038	0110017	469737	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

9.0 in air and 43.0 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522590

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33039	0110017	469813	Oxygen Sensor	
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

9.6 in air and 45.1 in oxygen. This sensor is nonlinear replaced under warranty with S/N 522591

Service Repair Sheet 67564
Contact Name Tracy Wang
Company/ Hospital Name Justec Shenzhen Co. Ltd
Direct Phone 0755-26068930/1/2
General Phone 0086 755 2606 8931
Opera Account 00006027
Email tracy.wang@justec.cn
Date Received 01/Sep/2020
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33040	0110017	469820	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

9.1 in air and 43.5 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522592

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33041	0110017	469879	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

8.7 in air and 41.7 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522593

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33042	0110017	470012	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Connector fault. Replaced under warranty with S/N 522594

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33043	0110017	470063	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Connector fault. Replaced under warranty with S/N 522595

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33044	0110017	470147	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

8.5 in air and 40.5 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522596

Service Repair Sheet 67564
 Contact Name Tracy Wang
 Company/ Hospital Name Justec Shenzhen Co. Ltd
 Direct Phone 0755-26068930/1/2
 General Phone 0086 755 2606 8931
 Opera Account 00006027
 Email tracy.wang@justec.cn
 Date Received 01/Sep/2020
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33045	0110017	470258	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

9.1 in air and 43.2 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522597

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33046	0110017	470275	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Connector fault. Replaced under warranty with S/N 522598

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33047	0110017	470551	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

8.8 in air and 42.3 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522599

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33048	0110017	471417	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

8.3 in air and 39.8 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522600

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33049	0110017	471436	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

Dead. Replaced under warranty with S/N 522601

Service Repair Sheet 67564
Contact Name Tracy Wang
Company/ Hospital Name Justec Shenzhen Co. Ltd
Direct Phone 0755-26068930/1/2
General Phone 0086 755 2606 8931
Opera Account 00006027
Email tracy.wang@justec.cn
Date Received 01/Sep/2020
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33050	0110017	474583	Oxygen Sensor	Return Stock / Demo Stock
				Time :0 Hour(s)

Parts Replaced

Qty: 1 0110017 Oxygen sensor R-17MED

8.0 in air and 37.9 in oxygen. This sensor is ok but replaced as a good will gesture with S/N 522602
