

Service Repair Sheet SRS67609

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Ashley Davies

Thames Medical Ltd

Office Manager

01903 522911

00011982

ashley@thamesmedical.com

PO: 01379 CG 10-10-2020

06/Oct/2020

Robert Connor

Viamed

Warranty

Notes 05/Oct/2020 Kate Griffiths

05/Oct/2020 Kate Griffiths

Customer would like to return RMA C02 Analyser SN: 230491. The capnograph sensor, fault light is illuminated red no green and display reads check sensor. Purchased January 2020.

06/Oct/2020 Robert Connor

1 x IRMA s/n 230491

Ready For quote

Repair Complete Signed

CC Connor 11-10-2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33146	Accessory	4420505	230491	Yes

4440010 x 1  
old S/N, SRS, SRN

4420505 x 1 FCC  
New S/N, SRS, SRN

4420512 x 1 FCC  
SRS, SRN

UPS x 1 FCC

Thames Medical Ltd  
 3 A'Becket Gardens  
 Worthing, West Sussex BN13 2BW  
 www.thamesmedical.com  
 email info@thamesmedical.com  
 tel 01903 522911  
 fax 01903 522922



# PURCHASE ORDER

order # PO-01379  
 date 05/10/2020

Experts in patient monitoring and anaesthesia

vendor address Viamed Ltd  
 15 Station Road  
 Cross Hills  
 Keighley, West Yorkshire  
 BD20 7DT

contact Kate  
 phone 01535 634542  
 fax 01535 635582

carrier  
 UPS

payment terms  
 Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SUB-TOTAL
TML-V-VM2500M1- 4420505RL2WR	VM2500M - IRMA C02 Analyser - Warranty Repair - Service Level 2 SN-230491	Warranty Repair of VM2500M - IRMA C02 Analyser	1	£0.00	£0.00
remarks	IRMA within warranty SRS67609				
				SUB-TOTAL	£0.00
				TAX	£0.00
				TOTAL	£0.00

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:  
 info@thamesmedical.com