

Service Repair Sheet SRS67558

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Tracy Wang

Company/ Hospital Name

Justec Shenzhen Co. Ltd

Department

Position

Direct Phone

General Phone

0086 755 2606 8931

Opera Account

00006027

Email

tracy.wang@justec.cn

Order Number

Date Received

01/Sep/2020

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 04/Aug/2020 Ryan Swaine

04/Aug/2020 Ryan Swaine

04/Aug/2020 Ryan Swaine

2 x R-23V sensors: V120669 - Failing to calibrate and V120668 - High output

01/Sep/2020 Robert Connor

V120669

V120668

Ready For quote

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33023	Oxygen Sensor	0110023	V120668	Yes
SRN33024	Oxygen Sensor	0110023	V120669	Yes

0140013 x 2 FCC.

Old S/N, SRN, SRS

0110023 x 2 FCC

New S/N, SRN, SRS

UPS x 1 FCC