

Service Repair Sheet SRS67406

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Hendrik J Kant

Royal Victoria Infirmary

Electronics and Medical Engineering Department

EME Support Technician

0191 2139792

00003785

Hendrik.Kant@nhs.net

06/Feb/2020

Robert Connor

Viamed

Quote

O/N

RTDN400039770

Notes 03/Feb/2020 Kate Griffiths

03/Feb/2020 Kate Griffiths

03/Feb/2020 Kate Griffiths

Hendrik has a faulty Tom Thumb SN: 0401430 "fault needle stuck at 18 cmH2o. Also requires a service"

06/Feb/2020 Robert Connor

Received 1 x tom thumb s/n 0401430

Ready For quote

CCNeeen 11.2.2020

Repair Complete Signed

CCNeeen 22.10.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32346	Tom Thumb	0310030	0401430	N

0380000 x ¹/_{0.25} @ £22.50 £90 as per Derek
S/N, SRS, SRN.

ZT 3785
ON SYSTEM
12 FEB 2020
QVM121870

0330203 x 1 @ £125.
SRS, SRN

UPS x 1 @ £10

Newcastle upon Tyne Hospitals NHS Trust

Electronics and Medical Engineering Department

Royal Victoria Infirmary

Queen Victoria Road

Newcastle upon Tyne

NE1 4LP

Email: Hendrik.Kant@nhs.net

Telephone: 0191 2139792

Please find enclosed equipment for repair

On completion please return to above address.

REPORTED FAULT

Enclosed Tom Thumb S/N: 0401430 for repair and calibration return no
SRS67406

Kind regards,

Hendrik J Kant