

Service Repair Sheet SRS67596

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Magnus Persson

Company/ Hospital Name

Swevet - AB

Department

Position

Direct Phone

46 416 258 21

General Phone

46 771215500

Opera Account

00006308

Email

magnus.persson@swevet.se

Order Number

Date Received

06/Oct/2020

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 22/Sep/2020 Zoey Teal

22/Sep/2020 Zoey Teal

Returning a VM-2500S. Serial number A1610160020.

Spo2 is not working and customer wants a rechargeable battery on the unit.

06/Oct/2020 Robert Connor

1 x VM2500S s/n A1610160020, with silicone cover.

Ready For quote

Repair Complete Signed

CCN 27-10-2020

| SRN | Equipment | Stock Ref | Serial Number | Warranty |
|----------|------------|-----------|---------------|----------|
| SRN33143 | Capnograph | 4410540 | A1610160020 | N |

4440000 x 1 FCC

S/N, SRS, SRN

NO fault found

UPS x 1 FCC