

Service Repair Sheet 67632

Contact Name	Saeed Namini
Company/ Hospital Name	St James University Hospital
Department	Medical Electronics
Position	Medical Electronics Technician
Direct Phone	0113 206 4829
General Phone	0113 243 3144
Opera Account	00002570
Email	Saeed.Namini@nhs.net
Date Received	02/Nov/2020
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33241	0110420	EJ35601069	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110420 Oxygen sensor MAX-14 (Fluke version)

10.6 in air and 49.0 in oxygen, this sensor is nonlinear and replaced under warranty with S/N FG64801176
