

Service Repair Sheet SRS67641

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600
A Graham

Sunderland Royal Hospital
Clinical Engineering

0191 5699933

00004992

16/Nov/2020

Catherine Spence

Viamed

Quote

Notes 16/Nov/2020 Catherine Spence

16/Nov/2020 Catherine Spence

16/Nov/2020 Catherine Spence

2 x tom thumb

S/N 9070 and VMI0011

a dial is sticking and due ppm and one is missing gauge face cover and due ppm

hospital job numbers 957621,957620

Ready For quote

CC, Neen 19.11.2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33244	Tom Thumb	0310030	VMC9070	N
SRN33245	Tom Thumb	0310030	VMI0011	N

0380000 x 2 @ £90 each.
S/N, SRS, SRN

0330203 x 2 @ £125 each.
SRS, SRN

0330212 x 2 @ £1.50 each.
SRS, SRN

UPS x 1 £10.

MEDICAL DEVICE RETURNS FORM

CLINICAL ENGINEERING
SUNDERLAND ROYAL HOSPITAL
KAYLL ROAD
SUNDERLAND
SR4 7TP

Tel: 0191 5699933
Fax: 0191 5699274

Date:

04/11/20

C.H.S. Asset No:

096307 + N/A.

The enclosed equipment is in need of repair.

Description: 2x Tom Thumb

C.H.S. Job No:

957621+957620

Make

Viamed

TNT Cons. No:

126994845

Model:

Tom Thumb

Order No: For Repair E/

Serial No:

9070 + VM20011

On Contract:

Estimate Requested YES/NO

Warranty:

Sent For Attention Of:

Viamed Ltd.
15 Station Road
Cross Hills
Keighley

Post Code

BD20 7DT.

Fault Description/Action to Take:

1x Missing Screw + due PPM
1x Dial Sticking + due PPM

Accessories Sent:

2x Tom Thumb

Sent By:

A. GRAHAM

Return To:

Clinical Engineering, Sunderland Royal Hospital, A Floor Theatre Block, Kayll Road,
Sunderland, SR4 7TP

Note: It is the responsibility of the receiver to confirm the contents that are listed above are correct.
Any damaged or missing items must be reported to the Electronics Department immediately.
DECLARATION OF CONTAMINATION STATUS OVERLEAF