



Zoey Teal <viamed.zoey.teal@gmail.com>

Fwd: Oxygen sensor for Philips Trilogy Evo2 messages

Steve Hardaker <office@viamed.co.uk>
Reply-To: steve.hardaker@viamed.co.uk
To: Zoey Teal <zoey.teal@viamed.co.uk>

Tue, Nov 24, 2020 at 6:10 PM

Sorry, forgot to cc you in.

Steve

----- Forwarded message -----

From: **Steve Hardaker** <steve.hardaker@viamed.co.uk>
Date: Tue, 24 Nov 2020 at 17:57
Subject: Oxygen sensor for Philips Trilogy Evo
To: <ashwin.andiapen@nhs.net>

Hi Ashwin,

Sorry for taking a while to respond to your query, I have only just received the information from MAXtec. The Max-125 sensor is made by Maxtec for Philips and we do not have access to that specific sensor, however, I have been advised that the Max-16 is compatible.

This is further backed up by the Philips Trilogy Evo accessories catalogue:

<https://www.philips.co.uk/healthcare/product/HCDS2110X11B/trilogy-evo-portable-hospital-to-home-ventilator/documentation>

On page 12, they list the sensor - [FiO2 sensor assembly \(1134668\)](#)



FiO2 sensor assembly (1134668)

The image clearly shows a Maxtec MAX-16 sensor, which we can supply @ £70.00+VAT each or £63.00+VAT each for 2+. Alternatively, we have a lower cost alternative in the R-45V, which is directly compatible to the Max-16 and is available as follows:

- P/n 0110045 - Oxygen sensor R-45V

1-off: £42.00+VAT

2-off: £38.00+VAT

3-10: £36.00+VAT

11+: £34.00+VAT

Carriage: FOC.

If you need a formal quotation, my colleague, Zoey will be able to help, or you can place an order by emailing orders@viamed.co.uk

Regards,

Steve Hardaker
Technical Support Manager
Viamed Ltd.

Please note: Viamed is now enacting a coronavirus contingency plan to allow sales and admin Staff to work from home, and I am now working remotely. Telephone calls to the main office will not be answered, but emails will. Orders may take longer to process but we aim to continue to operate and serve our customers as best we can.

Website: www.viamed.co.uk

email: steve.hardaker@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up

Twitter: twitter.com/ViamedLtd

Facebook: facebook.com/ViamedLtdUK

Company registered in England, No. 1291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

GDPR / Privacy Policy available at www.viamed.co.uk/policy

Steve Hardaker <office@viamed.co.uk>
Reply-To: steve.hardaker@viamed.co.uk
To: Zoey Teal <zoey.teal@viamed.co.uk>

Wed, Nov 25, 2020 at 10:13 AM

Hi Zoey,

Please can you issue a quote, thanks.

Regards,

Steve Hardaker
Technical Support Manager
Viamed Ltd.

Please note: Viamed is now enacting a coronavirus contingency plan to allow sales and admin Staff to work from home, and I am now working remotely. Telephone calls to the main office will not be answered, but emails will. Orders may take longer to process but we aim to continue to operate and serve our customers as best we can.

Website: www.viamed.co.uk

email: steve.hardaker@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up

Twitter: twitter.com/ViamedLtd

Facebook: facebook.com/ViamedLtdUK

Company registered in England, No. 1291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

GDPR / Privacy Policy available at www.viamed.co.uk/policy

----- Forwarded message -----

From: **ANDIAPEN, Ashwin (ST GEORGE'S UNIVERSITY HOSPITALS NHS FOUNDATION TRUST)**

<ashwin.andiapen@nhs.net>

Date: Wed, 25 Nov 2020 at 09:56

Subject: RE: Oxygen sensor for Philips Trilogy Evo

To: steve.hardaker@viamed.co.uk <steve.hardaker@viamed.co.uk>

Hi Steve,

Many thanks for looking into this for us.

Would it please be possible to have an official quote for 9 units of these?

We will need a quote to be able to place an order.

Kind Regards

Ashwin

From: Main Account [<mailto:viamedinbox@gmail.com>] **On Behalf Of** Steve Hardaker

Sent: 24 November 2020 17:57

To: ANDIAPEN, Ashwin (ST GEORGE'S UNIVERSITY HOSPITALS NHS FOUNDATION TRUST)
<ashwin.andiapen@nhs.net>

Subject: Oxygen sensor for Philips Trilogy Evo

This message originated from outside of NHSmail. Please do not click links or open attachments unless you recognise the sender and know the content is safe.

[Quoted text hidden]

This message may contain confidential information. If you are not the intended recipient please inform the sender that you have received the message in error before deleting it. Please do not disclose, copy or distribute information in this e-mail or take any action in relation to its contents. To do so is strictly prohibited and may be unlawful. Thank you for your co-operation.

NHSmail is the secure email and directory service available for all NHS staff in England and Scotland. NHSmail is approved for exchanging patient data and other sensitive information with NHSmail and other accredited email services.

For more information and to find out how you can switch, <https://portal.nhs.net/help/joiningnhsmail>