

Service Repair Sheet SRS67639

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Paul Simmonds

Leicester Royal Hospital

Medical Physics Department

0116 2585117

00002600

paul.simmonds@uhl-tr.nhs.uk

16/Nov/2020

Catherine Spence

Viamed

Quote

Notes 16/Nov/2020 Catherine Spence

16/Nov/2020 Catherine Spence

16/Nov/2020 Catherine Spence

Microstim battery no leads

M0004572

16/Nov/2020 Catherine Spence

has battery but no leads

Ready For quote

CGneen 26.11.2020

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33242	Microstim DB3	2510000	M0004572	N

2540010 x 1 @ £45

SRS,

SN:, SRS, SRN.

ups x 1 £10.

LEIC. ROYAL INFIRMARY.

TEL: 0116 2585117

Medical Physics Department

ALC 260

Date:

2/11/20

Dear Sirs,

I enclose

MICROSTIM NERVE STIMULATOR

- ☐ for warranty repair F.O.C.
- ☐ for recalibration
- ☐ for replacement
- ☐ a damaged

- ☒ for repair
- ☐ for overhaul
- ☐ for annual service

"NO OUTPUT"

Serial No.

M0004572

Our confirmation order (to follow) will be:

- ☐ Please quote for a replacement item, giving delivery dates.
- ☒ Please advise likely cost of repair.
- ☐ Should there be any additional cost please notify me and I will raise an order to cover that amount.

Please do not hesitate to contact me should there be any problem with its repair.

We look forward to its/their early return.

Yours faithfully

P. Simmonds

Electro Medical Service

MEDICAL PHYSICS DEPT.,
SANDRINGHAM BUILDING,
LEICESTER ROYAL INFIRMARY
LEICESTER. LEI 5HW

paul.simmonds@ukl-tr.nhs.uk