

Service Repair Sheet 67630

Contact Name	Khanh Thanh
Company/ Hospital Name	An An Import Export Co Ltd
Direct Phone	84 987 546 072
Opera Account	CID15101
Email	phukienthaythe@gmail.com
Date Received	27/Nov/2020
Booked in By	Catherine Spence

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33250	0110042	V102092	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0110042    Oxygen sensor R-42V

Dead, replaced under warranty with S/N V102163

---