

Service Repair Sheet SRS67457

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ivana Wancova

Company/ Hospital Name

Asqa A.S

Department

Position

Direct Phone

+420 272 761350

General Phone

Opera Account

00007356

Email

ivana.wancova@asqa.cz

Order Number

Date Received

23/Mar/2020

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 12/Mar/2020 Zoey Teal

12/Mar/2020 Zoey Teal

Sending in 7 VM2101 for repair.

A1803050012 - broken

A1603050058 â€" no values

51100324 - display

80100226 â€" does not work

70200175 - display

80800364 â€" no values

90300134 â€" no values

23/Mar/2020 Robert Connor

7 x VM2101 s/n A1803050012, A1803050058, 90300134, 80800364, 80100226, 70200175, 51200324

Ready For quote

Repair Complete Signed

CCN 1.12.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32442	Pulse Oximeter	0012101	A1803050012	Yes
SRN32443	Pulse Oximeter	0012101	A1803050058	Yes
SRN32444	Pulse Oximeter	0012101	90300134	2
SRN32445	Pulse Oximeter	0012101	80800364	2
SRN32446	Pulse Oximeter	0012101	80100226	2
SRN32447	Pulse Oximeter	0012101	70200175	2
SRN32448	Pulse Oximeter	0012101	51100324	2

0042101 x 7 FCC

SN, SRS, SEN

unable to repair

SRN32442 and SRN 32443 to be credited.

UPS

x

1

FCC