

EC DECLARATION OF CONFORMITY

**APPLICATION OF MEDICAL
DEVICE DIRECTIVE** : Annex II excluding section 4
93/42/EEC, as amended by Directive
2007/47/EC
RoHS Directive 2011/65/EU, along with
Annex IV, item 1 (b) exemption

**STANDARDS TO WHICH
CONFORMITY IS DECLARED** : EN 60601-1 3.1 Edition
EN60601-1-2; EN 14971
EN60601-1-6; EN ISO 80601-2-55

MANUFACTURER'S NAME : TELEDYNE INSTRUMENTS
ANALYTICAL INSTRUMENTS
a business unit of Teledyne Instruments, Inc.

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

PRODUCT NAME : Oxygen Analyzer


DEVICE CLASSIFICATION : Class IIa (Annex IX, Rule 10)


MODEL NUMBER : AX300 and AX300-I

Authorized Representative : Viamed Limited
15 Station Road, Cross Hills Kieghley.
W. Yorkshire BD20 7DT United Kingdom
Ph: + 44 (0) 1535 634542
URL: www.viamed.co.uk

Registration Information : NB: BSI Netherlands /# 2797
CE Certification # CE 02000

We, Teledyne Analytical Instruments, a business unit of Teledyne Instruments, Inc, declare that the above mentioned products meet the provision of the Council Directive 93/42/EEC, as amended by Council Directive 2007/47/EC, for Medical Devices.

SIGNATURE: 
FULL NAME: Roger Starlin
POSITION: Manager, Quality Assurance
Date: 9-12-19

SIGNATURE: 
FULL NAME: Angel Alegria
POSITION: New Products Manager
Date: 9-12-19