

Service Repair Sheet 67664

Contact Name	Alan Cook
Company/ Hospital Name	V-Tech UK Ltd
General Phone	0208 498 1288
Opera Account	00012679
Email	Alan@v-techuk.com
Date Received	10/Dec/2020
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33272	0110122	536303	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110122 Oxygen sensor - Automotive R-22A

11.3 in air and 54.6 in oxygen, this sensor is nonlinear and replaced under warranty with S/N 554642
