

Service Repair Sheet SRS67578

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Tracy Wang

Company/ Hospital Name

Justec Shenzhen Co. Ltd

Department

Position

Direct Phone

General Phone

0086 755 2606 8931

Opera Account

00006027

Email

tracy.wang@justec.cn

Order Number

Date Received

07/Dec/2020

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 07/Sep/2020 Ryan Swaine

07/Sep/2020 Ryan Swaine

07/Sep/2020 Ryan Swaine

R-22MED SN 457482 - Low Output

In Warranty

07/Dec/2020 Catherine Spence

received in

Ready For quote

Repair Complete Signed

C. Spence 14.12.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33270	Oxygen Sensor	0110022	457482	Yes

~~get~~
0140013 x 1 @ £0.
old SN, SRS, SRN

0110022 x 1 @ £0.
SN, SRS, SRN.

UPS Exp Saver £0.
check to see if they want
to ship with next order.