

Service Repair Sheet 67578

Contact Name	Tracy Wang
Company/ Hospital Name	Justec Shenzhen Co. Ltd
General Phone	0086 755 2606 8931
Opera Account	00006027
Email	tracy.wang@justec.cn
Date Received	07/Dec/2020
Booked in By	Catherine Spence

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33270	0110022	457482	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110022 Oxygen sensor R-22MED

9.2 in air and 44.5 in oxygen, this sensor is nonlinear and replaced under warranty with S/N 532029
