

Service Repair Sheet SRS67638

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Date Promised to Return: 01 Jan 1970 -3600
Roger Smith
Leicester Royal Infirmary
Neonatal Unit
Senior Critical Care Technologist
0116 258 6487

00002600
roger.smith@uhl-tr.nhs.uk

14/Dec/2020
Robert Connor
Viamed
For Servicing

Notes 13/Nov/2020 Steve Hardaker

13/Nov/2020 Steve Hardaker

13/Nov/2020 Steve Hardaker

Returning 3x TT480 Tom Thumbs for annual servicing.

Serial numbers: 040408, 040409 and 040967.

Note: 3 Service Exchange units (SE006,7,8) sent in advance as loan units on SOR927 whilst these are back with us. We previously capped the pressure to 40cmH2O at their request but Roger Smith on 16/5/18 advised "Yes, it is acceptable to set the new PIP default to 42.1 +/- 1 mbar as per latest standard QA spec."

01/Dec/2020 Steve Hardaker

Loan units have been received by the Trust. When completed, please send service reports to:

- UHLEquipmentMaintenanceContracts@uhl-tr.nhs.uk

- sarah.fletcher@uhl-tr.nhs.uk

10/Dec/2020 Steve Hardaker

13/Nov/2020 Steve Hardaker

13/Nov/2020 Steve Hardaker

13/Nov/2020 Steve Hardaker

Returning 3x TT480 Tom Thumbs for annual servicing.

Serial numbers: 040408, 040409 and 040967.

Note: 3 Service Exchange units (SE006,7,8) sent in advance as loan units on SOR927 whilst these are back with us. We previously capped the pressure to 40cmH2O at their request but Roger Smith on 16/5/18 advised "Yes, it is acceptable to set the new PIP default to 42.1 +/- 1 mbar as per latest standard QA spec."

01/Dec/2020 Steve Hardaker

Loan units have been received by the Trust. When completed, please send service reports to:

- UHLEquipmentMaintenanceContracts@uhl-tr.nhs.uk

- sarah.fletcher@uhl-tr.nhs.uk

Shipped to us 10/12/20 by special delivery - NY0077-38419GB.

14/Dec/2020 Robert Connor

3 x TT480 s/n 040408, 040409, 040967

Ready For quote

Repair Complete Signed

CGN 16.12.2020

update
CVM21570

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33290	Tom Thumb	0310030	040408	2
SRN33291	Tom Thumb	0310030	040409	2
SRN33292	Tom Thumb	0310097	040967	2

03800000 x
S/N SRN, SLS

3 £90 each

UPS

x

1

£10