

## EMRTS Cymru - Electromed Fault Report

Database ID: -MOKee1PFPpeLVg9RXQ3



### Fault Report

**ID:** -MOKee1PFPpeLVg9RXQ3

**Date:** 12/12/20 - 06:33

**Base:** H67 - Cardiff

**Reported by:** Darragh Coghlan

### Location:

Device was in Viamed Module 02 - Barcode: 01524

**Electromed ID:** 01624

**Manufacturer:** Viamed

**Model:** VM-2500-S

**Item:** 01624

### Fault Details

**Damaged/Mechanical:** YES

**Software Issue:** NO

### Fault Description:

When conducting a full bag check. I noted the screen is not fully visible and is missing pixels. This is making it very hard to read and give accurate observations. I will discuss with day team and Jay hughes whether to leave it in service or not

### Case Details

**Near Miss:** NO

**Patient Harm:** NO