

Service Repair Sheet SRS67662

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Magnus Persson

Swevet - AB

46 416 258 21

46 771215500

00006308

magnus.persson@swevet.se

10/Dec/2020

Robert Connor

Viamed

Quote

Notes 02/Dec/2020 Zoey Teal

02/Dec/2020 Zoey Teal

Sending in a AX300-I. The customer can not calibrate it. Sensor 0.00 and the screen is flashing.

Serial number 32848 but this may be wrong

10/Dec/2020 Robert Connor

1 x AX300 s/n 328348, with stand, cable s/n 0014010, R-17MED s/n 135325, flow diverter and 3 x AA battery

Ready For quote

Repair Complete Signed

CC, Neen 14.12.2020

CC, Neen 15.12.2020

CC, Neen 23.12.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33278	Oxygen Monitor	0111230	328348	N
SRN33279	Oxygen Monitor Cable	0131202	0014010	N
SRN33280	Oxygen Sensor	0110017	135325	N

0140004 x 1 €70 - half an hour = €45
 SIN, SRS, SRN

~~0140025 x 1~~
~~SIN, SRS, SRN~~

0131214 x 1 €2.88
 SRS, SRN

→ Add

0140025 x 1 = €0.
 SIN, SRS, SRN BIC ?

0140012 x 1 €0.00
 Old SIN, SRN, SRS INS ?

~~0140012 x 1~~
~~Old SIN, SRN, SRS~~ €0.00

0110017 x 1 x 1 €32.00
 SIN, SRS, SRN

... € To Add.

Service Repair Sheet SRS67662

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Magnus Persson

Swevet - AB

46 416 258 21

46 771215500

00006308

magnus.persson@swevet.se

10/Dec/2020

Robert Connor

Viamed

Quote

Notes 02/Dec/2020 Zoey Teal

02/Dec/2020 Zoey Teal

Sending in a AX300-I. The customer can not calibrate it. Sensor 0.00 and the screen is flashing.

Serial number 32848 but this may be wrong

10/Dec/2020 Robert Connor

1 x AX300 s/n 328348, with stand, cable s/n 0014010, R-17MED s/n 135325, flow diverter and 3 x AA battery

Ready For quote

Repair Complete Signed

CC Neen 14.12.2020 CC Neen 15.12.2020

CC Neen 23.12.2020

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33278	Oxygen Monitor	0111230	328348	N
SRN33279	Oxygen Monitor Cable	0131202	0014010	N
SRN33280	Oxygen Sensor	0110017	135325	N

0140004 x 1 €70 - half an hour = €45
 SIN, SRS, SRN

~~0140025 x 1~~
~~SIN, SRS, SRN~~

0131214 x 1 €2.88
 SRS, SRN, SIN

→ Add

0140025 x 1 = €0.
 SIN, SRS, SRN BIC ?

0140012 x 1 €0.00
 Old SIN, SRN, SRS INS ?

~~0140012 x 1~~ €0.00
~~Old SIN, SRN, SRS~~

0110017 x 1 x 1 €32.00
 SN, SRS, SRN

UPS x 1 - 32 x 24 x 10cm € To Add.