



Sophie Lines &lt;viamed.sophie.lines@gmail.com&gt;

---

**Fwd: Sensors**

1 message

---

**Main Account** <office@viamed.co.uk>  
To: Sophie Lines <sophie.lines@viamed.co.uk>

Tue, Dec 29, 2020 at 9:52 AM

----- Forwarded message -----

From: **BLACKWELL, David (DARTFORD AND GRAVESHAM NHS TRUST)** <davidblackwell@nhs.net>

Date: Tue, 29 Dec 2020 at 09:47

Subject: Sensors

To: [enquiries@viamed.co.uk](mailto:enquiries@viamed.co.uk) <[enquiries@viamed.co.uk](mailto:enquiries@viamed.co.uk)>Cc: TILLMAN, Michael (DARTFORD AND GRAVESHAM NHS TRUST) <[michael.tillman@nhs.net](mailto:michael.tillman@nhs.net)>

Good morning

We ordered 2x 0110435 oxygen sensors for the Medtronic 980 Ventilators.

When installed, these failed, both were 16% out on all tests.

Both were opened as we were desperate to get the ventilator out, unfortunately we had to use the old cell.

Could this be looked into please/refunded.

Thank you.

**Dave Blackwell****Medical Engineering Technician Supervisor*****EBME Department******Darent Valley Hospital - Darenth Wood Road, Dartford, Kent .DA2 8DA***<http://www.dvh.nhs.uk/>**Gravesham Community Hospital - Bath St, Gravesend DA11 0DG*****Queen Mary's Hospital - Frogna Avenue, Sidcup, Kent. DA14 6LT.***

<http://www.qmh.oxleas.nhs.uk/>



\*\*\*\*\*  
\*\*\*\*\*

This message may contain confidential information. If you are not the intended recipient please inform the sender that you have received the message in error before deleting it. Please do not disclose, copy or distribute information in this e-mail or take any action in relation to its contents. To do so is strictly prohibited and may be unlawful. Thank you for your co-operation.

NHSmial is the secure email and directory service available for all NHS staff in England and Scotland. NHSmial is approved for exchanging patient data and other sensitive information with NHSmial and other accredited email services.

For more information and to find out how you can switch, <https://portal.nhs.net/help/joiningnhsmial>