

Service Repair Sheet SRS67544

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Dieter Eckert

DEM Medizintechnik

49 519397590

00007016

info@dem-med.de

18/Jul/2020

Catherine Spence

Viamed

Quote

Notes 18/Jul/2020 Catherine Spence

18/Jul/2020 Catherine Spence

18/Jul/2020 Catherine Spence

Microstim and leads

S/N M0004318

Divice has loose contact, works for a few seconds after inserting new batteries and then stops working

Ready For quote

CGreen 18-8-2020

Repair Complete Signed

CGreen 4-1-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32946	Microstim DB3	2510000	M0004318	N

2540010 x 1 @ € ~~49.42~~ \$ 57.29
S/N, SRS, SRN

UPS x 1 @ £ \$
32x26x12cm

Sarataweg 1a
D-29640 Schneverdingen

+49 5193 9759-0
+49 5193 9759-11
info@dem-med.de
www.dem-med.de

DEM Medizintechnik | Sarataweg 1A | D-29640 Schneverdingen
Electromedical Equipment
VIAMED
15 Station Road, Cross Hills
BD207DT Keighley, West Yorkshire
England

Rücksendung Nr. 20024

Seite: 1
Lief.-Nr.: 1532
Bearbeiter: Yvonne Fischer
Steuernr.: 41 111 01207
USt-IdNr.: DE171270984
Datum: 07.07.2020

Pos	Menge	Art.-Nr	Text	Einzelpreis EUR	Gesamtpreis EUR
1	1	Microstim DB3	Error description: Device has a loose contact. It works only a few seconds after inserting the new battery and no longer works Supramaximaler Nervenstimulator incl. Patientcable SN: M0004318		
Gesamt Netto					0,00
steuerfrei				0,00	
Gesamtbetrag					0,00

Mit freundlichen Grüßen

Dieter Eckert
DEM Medizintechnik