

Service Repair Sheet SRS67702

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical Ltd

Department

Position

Office Manager

Direct Phone

01903 522911

General Phone

Opera Account

00011982

Email

ashley@thamesmedical.com

Order Number

PO-01465

Date Received

22/Jan/2021

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 19/Jan/2021 Kate Griffiths

19/Jan/2021 Kate Griffiths

Faulty VM-2500-M SN:80230138

"Customer is reporting pins being broken on the charging port. Please could you send a quote before work takes place please"

22/Jan/2021 Catherine Spence

Capnograph S/N 80230138 with mounting and charger. Pin stuck in charger connector

Ready For quote

CGNeen 27.1.2021

Repair Complete Signed

CGNeen 28.1.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33341	Capnograph	4410530	80230138	N

DIST

4440000
S/N, SRS, SRN

£90 per hr
= £45
0.5 hr @

4400080 x 1
SRS, SRN


£42.76

4430065 x 1
SRS, SRN

£12.54

UPS x 1 @ £10.

Thames Medical Ltd
8 A'Becket Gardens
Worthing, West Sussex BN13 2BW
www.thamesmedical.com
email info@thamesmedical.com
tel 01903 522911
fax 01903 522922
Experts in patient monitoring and anaesthesia



PURCHASE ORDER

order # PO-01465
date 19/01/2021

vendor address **Viamed Ltd**
15 Station Road
Cross Hills
Keighley, West Yorkshire
BD20 7DT

contact Kate
phone 01535 634542
fax 01535 635582

carrier UPS
payment terms Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SUB-TOTAL
TML-V-VM2500M1- RL2WR	VM2500M - Warranty Repair - Service Level 2 SN:80230138	Warranty Repair of VM2500M Mainstream Capnograph	1	£0.00	£0.00
remarks	SRS67702 Please quote before repair.			SUB-TOTAL TAX TOTAL	£0.00 £0.00 £0.00

*Not Warranty
Sold 17-9-2014
CG 22-1-2021*

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:
info@thamesmedical.com