Service Repair Sheet SRS67668

Contact Name

Company/ Hospital Name

Department

Position

**Direct Phone** 

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lewis Ball

Inspiration Healthcare Ltd

Service Department

Senior Service Engineer

07584 147057

01455 840555

00012240

lewis.ball@inspiration-healthcare.com

22/Jan/2021

Catherine Spence

Viamed

Warranty

Notes 07/Dec/2020 Steve Hardaker

07/Dec/2020 Steve Hardaker

07/Dec/2020 Steve Hardaker

Returning R-47V s/n V104325 under warranty.

Failed in a vent due to a very high reading on one of the cathodes.

18/Jan/2021 Steve Hardaker

07/Dec/2020 Steve Hardaker

07/Dec/2020 Steve Hardaker

07/Dec/2020 Steve Hardaker

Returning R-47V s/n V104325 under warranty.

Failed in a vent due to a very high reading on one of the cathodes.

- Chased 18/1/21 - he will return it this week.

22/Jan/2021 Catherine Spence

received in

Ready For quote

Repair Complete Signed

<u>CGNEON 1-12-</u>2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33340	Oxygen Sensor		V104325	Yes -

0140013 x 1 @ 60.

NEW SIN, SES, SEN

UPS X 1 0 ED.

## Blott

From:

Lewis Ball

Sent:

19 January 2021 10:46

To:

Subject:

Warehouse

FW: Faulty R-47V V104325

Please print off and I will give you the cell to send to Viamed.

Regards

## Lewis Ball

Senior Service Engineer

## Inspiration Healthcare

Gildor House, West Street, Earl Shilton, Leicester, Leicestershire LE9 7EJ

T 01455 840555

M 07584 147057

E lewis.ball@inspiration-healthcare.com

W inspiration-healthcare.com

24 hour emergency hire and support call +44 (0) 1455 840 555





Winners of the Queen's Awards for Enterprise: International Trade 2019







This email and any attachment are intended solely for the addressee, are strictly confidential and may be legally privileged. If you are not the intended recipient any reading, dissemination, copying or any other use or reliance is prohibited. If you have received this email in error please notify the sender immediately by email and then permanently delete the email. Copyright reserved. All communications, incoming and outgoing, may be recorded and are monitored for legitimate business purposes. The security and reliability of email transmission cannot be guaranteed. It is the recipient's responsibility to scan this e-mail and any attachment for the presence of viruses. Inspiration Healthcare Group pic and its subsidiaries exclude all liability for any loss or damage whatsoever arising or resulting from the receipt, use or transmission of this email. Any views or opinions expressed in this email are those of the author only.

From: Steve Hardaker <viamed.steve.hardaker@gmail.com>

Sent: 14 January 2021 17:21

To: Lewis Ball < Lewis. Ball@inspiration-healthcare.com>; Mark Radbourne < Mark. Radbourne@inspiration-

healthcare.com>

Subject: Fwd: Faulty R-47V V104325

Hi Lewis,