

Service Repair Sheet 67691

Contact Name	Alan Marples
Company/ Hospital Name	Chesterfield Royal Hospital
Department	Department of Clinical Engineering
Position	Specialist Clinical Engineer
Direct Phone	01246 513065
General Phone	01246 277 271
Opera Account	00001170
Email	alanmarples@nhs.net
Date Received	27/Jan/2021
Booked in By	Catherine Spence

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33353	0110017	565706	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0110017    Oxygen sensor R-17MED

14.1 in air and 67.8 in oxygen. This sensor is high output and replaced under warranty with S/N 566033

---