

Calibration Verification

The UltraMaxO2 Oxygen Analyser has been tested for accuracy using medical grade Oxygen and air supplied by BOC. The output values were measured in accordance with procedure: VM3COP43.14.

Medical test gases used:

Medical Oxygen (BOC)	PL 0735/5000
Purity	99.5% (min)
Batch No.	12061018

Medical Air (BOC)	PL 0735/5002R
Purity	21.0%
Batch No.	11702744

Serial number of the UltraMaxO2	EB27311036
REF	R221P11-001

Oxygen input:	UltraMaxO2 reading:
99.5%	99.3%

Air input	UltraMaxO2 reading:
21.0%	20.7%

Air Flow Test:	Pass
Pressure Test:	Pass

Tested by:

Signed: