

Service Repair Sheet SRS67707

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Rennie Saverton

Inspiration Healthcare Ltd

Field Service Engineer

01455 840555

00012240

Rennie.Saverton@inspiration-healthcare.com

03/Feb/2021

Catherine Spence

Viamed

Warranty

Notes 22/Jan/2021 Steve Hardaker

22/Jan/2021 Steve Hardaker

22/Jan/2021 Steve Hardaker

Returning 4 Viamed Oxygen sensors that have failed prematurely whilst in use.

3x R-47V

V104251 (manufactured 2020-03)

V104575 (manufactured 2020-10)

V104578 (manufactured 2020-10).

1x R-22MEDV

V114652 (manufactured 2020-04).

"The sensors will be returned to you for investigation and warranty replacement, please use our reference CC959 in all correspondence regarding this matter."

03/Feb/2021 Catherine Spence
received in

Ready For quote

Repair Complete Signed

CGNEEN 10-2-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33368	Oxygen Sensor	0110047	V104251	Yes
SRN33369	Oxygen Sensor	0110047	V104575	Yes
SRN33370	Oxygen Sensor	0110047	V104578	Yes
SRN33372	Oxygen Sensor	0110021	V114652	Yes

0140013 x 3 FOC
S/N'S, SRS, SRN → SRN33368-70.

0110047 x 3 FOC
NEW S/N, SRS, SRN

0140011 x 1 FOC
S/N, SRS, SRN - 33372