

Service Repair Sheet SRS67732

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Mike Brampton

Company/ Hospital Name

Thames Medical

Department

Position

Direct Phone

General Phone

01903 522911

Opera Account

0011982

Email

Order Number

Date Received

12/Feb/2021

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Warranty

Notes 12/Feb/2021 Catherine Spence

12/Feb/2021 Catherine Spence

12/Feb/2021 Catherine Spence

VM2160 and ear clip

S/N 50001690 and clip EBD00431

last in on SRS67693 for a upgrade screen no accessories

Vet Probe supplied not working on unit

emails sent between mike and MG to check firmware. please treat as urgent

Ready For quote

Repair Complete Signed

C. Green 12.2.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33386	Pulse Oximeter	0012175	50001690	Yes
SRN33387	Pulse Oximetry Ear Clip Sensor	0015000	EBD00431	N

0042160 x 1 FOC
S/N, SRS, SRN

0040001 x 1 FOC
S/N, SRS, SRN

UPS x 1 FOC

Thames Medical Ltd
 8 A'Becket Gardens
 Worthing, West Sussex BN13 2BW
 www.thamesmedical.com
 email info@thamesmedical.com
 tel 01903 522911
 fax 01903 522922
 Experts in patient monitoring and anaesthesia



PURCHASE ORDER

order # PO-01486
 date 10/02/2021

vendor address Viamed Ltd
 15 Station Road
 Cross Hills
 Keighley, West Yorkshire
 BD20 7DT

contact Kate
 phone 01535 634542
 fax 01535 635582

carrier
 UPS

payment terms
 Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SUB-TOTAL
TML-V-VM2160RL2 WR	VM2160 - Warranty Repair - Service Level 2	Warranty Repair of VM2160 Pulse Oximeter	1	£0.00	£0.00
remarks	Unit displaying sensor fault with any clip sensors old or new.				SUB-TOTAL £0.00
	UCL				TAX £0.00
					TOTAL £0.00

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:
 info@thamesmedical.com