

Service Repair Sheet 67728

Contact Name	Neil Flanagan
Company/ Hospital Name	Spectrum Medical
Position	Procurement Specialist
Direct Phone	01242 650120
General Phone	01242 650120
Opera Account	00012812
Email	Neil.Flanagan@spectrummedical.com
Date Received	17/Feb/2021
Booked in By	Catherine Spence

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33425	0110021	V115391	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110021 Oxygen sensor R-22MEDV

10.3 in air and 47.5 in oxygen but leaking. Replaced under warranty with S/N V115946
