

0110017 R17Med
£42.00

Service Repair Sheet SRS67690

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Romeo Chekem / Martin Houghton

Kettering General Hospital

Medical Devices Department

Medical Electronics Supervisor

01536 492269

00002380

romeo.chekem@nhs.net

14/Jan/2021

Catherine Spence

Viamed

Quote

emails sent x2
by kate.

Notes 06/Jan/2021 Kate Griffiths

06/Jan/2021 Kate Griffiths

Customer has a faulty AX300, 0111230.

11/Jan/2021 Kate Griffiths

06/Jan/2021 Customer has a faulty AX300, 0111230.

"The fault is: not reading air". S/n 301626

PO S113513 rec'd and uploaded.

14/Jan/2021 Catherine Spence

received in with cable and stand

Telephoned
customer myself
cg 18-2-2021
replace
Add R17Med to
Order and proceed
cg 18-2-21

Ready For quote

Repair Complete Signed

CGNEEN 17-2-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33335	Oxygen Monitor	0111230	301626	N
SRN33336	Oxygen Monitor Cable	0131202	0003838	N

0140004 x 1 @ £0.

S/N, SRS, SEN.

NO fault found

0140025 x 1 @ £10

S/N, SRS, SEN.

NO fault found.

0110017 x 1 £42 → Added as agreed
by engineer. cg 19-2-2021

S/N, SRS, SEN.

NO fault found.

PURCHASE ORDER

Kettering General Hospital
NHS Foundation Trust

Purchase Order No : S113513

REGISTER FOR NEW SOURCING OPPORTUNITIES

HTTPS://WWW.SUPPLYING2NHS.COM

PURCHASING AND SUPPLIES DEPARTMENT
KETTERING GENERAL HOSPITAL
ROTHWELL ROAD
KETTERING
NORTHANTS NN16 8UZ
kgh-tr.supplies@mailbox@nhs.net

Y41096 MEDICAL ELECTRONICS DEPARTMENT

Supplier	
VIAMED 15 STATION ROAD CROSS HILLS KEIGHLEY WEST YORKSHIRE BD20 7DT	
Fax	
Email	sales@viamed.co.uk

Deliver To
RECEIPT & DISTRIBUTION POINT KETTERING GENERAL HOSPITAL ROTHWELL ROAD KETTERING NORTHANTS NN16 8UZ

Invoice To	
FINANCIAL SERVICES DEPARTMENT KETTERING GENERAL HOSPITAL ROTHWELL ROAD KETTERING NORTHANTS NN16 8UZ	
Electronic Invoices to : kgh-tr.finance@nhs.net	
Date	11/01/21
Internal	R121410
	Page 1

Line	Product / Service Description	Quantity	Price	Total Line Value
1	REPAIR FAULT NOT READING AIR Teledyne Analytical Instruments Model AX300-1 SN:301626	1.00	90.00	90.00
2	CARRIAGE	1.00	10.00	10.00
STANDARD TERMS AND CONDITIONS APPLY FOR A COPY OF THE LATEST TERMS PLEASE VISIT https://www.gov.uk/government/publications/nhs-standard-terms-and-conditions-of-contract-for-the-purchase-of-goods-and-supply-of-services				Total Less Tax £100.00
				Total VAT £20.00
				Total Order Value £120.00