

0116017 R17med  
£42.00

Service Repair Sheet SRS67690

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Romeo Chekem / Martin Houghton

Company/ Hospital Name

Kettering General Hospital

Department

Medical Devices Department

Position

Medical Electronics Supervisor

Direct Phone

01536 492269

General Phone

Opera Account

00002380

Email

romeo.chekem@nhs.net

Order Number

Date Received

14/Jan/2021

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

emails sent x2  
by kate.

Notes 06/Jan/2021 Kate Griffiths

06/Jan/2021 Kate Griffiths

Customer has a faulty AX300, 0111230.

11/Jan/2021 Kate Griffiths

06/Jan/2021 Customer has a faulty AX300, 0111230.

"The fault is: not reading air". S/n 301626

PO S113513 rec d and uploaded.

14/Jan/2021 Catherine Spence

received in with cable and stand

Telephoned  
Customer myself  
cg 18-2-2021  
Replace  
Add R17med to  
Order and proceed  
cg 18-2-21

Ready For quote

Repair Complete Signed

CGREEN 17-2-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33335	Oxygen Monitor	0111230	301626	N
SRN33336	Oxygen Monitor Cable	0131202	0003838	N

0140004 x 1 @ £0.

S/N, SRS, SEN.

NO fault found

0140025 x 1 @ £10 £0. cg 22-2-2021

S/N, SRS, SEN.

NO fault found.

0110017 x 1 £42 → Added as agreed  
by engineer. cg 19-2-2021

S/N, 571950

1 SRS x 1 f.D.