

Service Repair Sheet SRS67723

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Seri Cooke

Emergency Medical Retrieval & Transfer Service

0300 3000 057

00012969

Seri.Cooke@wales.nhs.uk

16/Feb/2021

Catherine Spence

Viamed

Quote

Notes 08/Feb/2021 Zoey Teal

08/Feb/2021 Zoey Teal

Sending in a Capnograph VM-2500-S. Serial number A1608160004. The screen is damaged and the pixels are intermittent.

16/Feb/2021 Catherine Spence
received in with silicone cover

Ready For quote

CANON 17-2-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33418	Capnograph	4410520	A1608160004	N

4440000 x half an hr @ £90 per hour = £45.00
S/N, SRS, SRN.

4430120 x 1 @ £72.00.
S/N, SRS, SRN

UPS x 1 £10.

EMRTS Cymru - Electromed Fault Report
Database ID: -MT0cgkgUfnAz4Sv-8oM

Fault Report

ID: -MT0cgkgUfnAz4Sv-8oM

Date: 08/02/21 - 12:32

Base: H57 - Dafen

Reported by: Jason Hughes

Location:

Location Unknown

Electromed ID: 04218

Manufacturer: Viamed

Model: VM-2500-S

Item: 04218

Fault Details

Damaged/Mechanical: YES

Software Issue: NO

Fault Description:

Screen damaged. Pixels are intermittent.

Case Details

EPR #: N/a

Near Miss: NO

Patient Harm: NO