



Calibration Certificate

The VN202 Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

Calibrated voltage source	Microcal
ISO calibration index number	CE076
Serial number of the Microcal	6939K9
Certificate number of the Microcal	1646861
Serial number of the VN202	02444
Certificate number of the VN202	1614682195

Voltage input:	Target reading:	VN202 reading:
4.78mV	10%	9.9%
9.99mV	20.9%	20.8%
10.0mV	21%	20.9%
11.95mV	25%	25.0%
23.9mV	50%	50.0%
35.85	75%	75.0%
47.8mV	100%	100.0%

Tested by: M.Green

Date: 02/03/2021

Signed: