

Service Repair Sheet SRS67781

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Jan Bowen

Company/ Hospital Name

Huntleigh Healthcare

Department

Diagnostic Products Division

Position

Direct Phone

02920 447097

General Phone

Opera Account

00011875

Email

jan.bowen@arjo.com

Order Number

Date Received

26/Mar/2021

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Notes 25/Mar/2021 Sarah Walton

25/Mar/2021 Sarah Walton

Faulty V1000 unit.

S/N: PR01883A16

26/Mar/2021 Catherine Spence

Received with batteries

It is crackling and very low audio

Ready For quote

CGreen 29.3.2021

Repair Complete Signed

CGreen 1-4.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33496	Foetal Simulator	1410000	PR01883A16	N

NOT SURE IF
DISTRIBUTED

if so then

£38.25

14800000 x 1 £45
S/N, SRS, SRN

1430309 x 1 £0.
SRS, SRN

UPS x 1 £10.

WITH COMPLIMENTS

HUNTLEIGH

Hi,

The Fault with the V1000 Foetal Heart Simulator is that it is crackling and very low Audio.

Regards,

Jan Bowden


Huntleigh Healthcare Ltd • Diagnostic Products Division • 35 Portmann Road • Cardiff • CF24 5HN • United Kingdom
Tel: +44 (0)29 2048 5885 • Fax: +44 (0)29 2049 2520 • Mail: sales@huntleigh-diagnostics.co.uk
www.huntleigh-diagnostics.com • Registered No: 942245 • Registered Office: 310-312 Dallow Road, Luton, LU1 1TD

AN ARJONHUNTLEIGH
MEMBER OF THE GETINK

Issues 13 Unread 15 ISO Tasks | H/W: KG Future

Active Users

Search



meetings



opera



producti



returns



sales



stock



stockboo



x

ct Currency

ds| Euros| Usd

Values in Pounds @ 1: 1.00

410000			
Reference	Quantity	Unit Cost	Total Line
480000	1	45.00	45.00
430309	1	15.00	15.00
		Parts Cost	60.00

BEFORE SENDING QUOTE

Quote Sent