

Service Repair Sheet SRS67753

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Rehab Attia

Royal Victoria Infirmary Hospital

Medical Electronics Technician

0191 2839794

00003785

rehab.attia@nhs.net

14/Mar/2021

Catherine Spence

Viamed

Quote

Notes 05/Mar/2021 Kate Griffiths

05/Mar/2021 Kate Griffiths

Customer has sent in Tom Thumb for "repair/annual service". s/n: 040179. They sent this via TNT before they asked for an SRS no. TNT: 535082940.

14/Mar/2021 Catherine Spence

received in

Giving unstable reading at maximum adjustable pressure and unable to calibrate

Ready For quote

CG Neen 15.3.2021

Repair Complete Signed

CG Neen 6.4.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33461	Tom Thumb	0310030	040179	N

0380000 x 1 @ £90

SIN, SRS, SRN-

0330203 x 1 @ £125

SRS, SRN.

0330212 x 1 @ £+50 2.00.

SRS, SRN

UPS x 1 £10