

Service Repair Sheet SRS67786

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Mike Brampton

Thames Medical

Managing Director

01903 522911

00011982

mike@thamesmedical.com

06/Apr/2021

Robert Connor

Viamed

Warranty

Notes 30/Mar/2021 Kate Griffiths

30/Mar/2021 Kate Griffiths

Customer says "we have just had a lingual clip sent back to us (intermittent working) ie it works unless the jaws are opened wide. . ."

S/n: ESKB 01686

Bought December 2020.

06/Apr/2021 Robert Connor

0015000*1 s/n ESKB01686

Ready For quote

Repair Complete Signed

Connor 6.4.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33515	Pulse Oximetry Ear Clip Sensor	0015000	ESKB01686	Yes

0040003 x 1 @ £0
old S/N, SRS.

0015000 x 1 @ £0.
S/N, SRS, SRN

UPS x 1 @ £0.

Thames Medical Ltd
8 A Becket Gardens
Worthing, West Sussex BN13 2BW
www.thamesmedical.com
email info@thamesmedical.com
tel 01903 522911
fax 01903 522922



PURCHASE ORDER

Experts in patient monitoring and anaesthesia

order # PO-01523
date 30/03/2021

vendor address **Viamed Ltd**
15 Station Road
Cross Hills
Keighley, West Yorkshire
BD20 7DT

contact Kate
phone 01535 634542
fax 01535 635582

carrier payment terms
UPS Net 30

ITEM	DESCRIPTION	QUANTITY	UNIT PRICE	SUB-TOTAL
TML-MISC	Clip Sensor Warranty Repair - SRS67786	1	£0.00	£0.00
			SUB-TOTAL	£0.00
			TAX	£0.00
			TOTAL	£0.00

Please acknowledge this order and advise the delivery date by fax: +44 (0) 1903 522922 and/or by email:
info@thamesmedical.com