

Service Repair Sheet 67786

Contact Name	Mike Brampton
Company/ Hospital Name	Thames Medical
Position	Managing Director
General Phone	01903 522911
Opera Account	00011982
Email	mike@thamesmedical.com
Date Received	06/Apr/2021
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33515	0015000	ESKB01686	Pulse Oximetry Ear Clip Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0015000    Pulse oximetry Lingual Sensor

The SpO2 lingual sensor has been replaced under warranty with SN:ESKB01690.

---