

Service Repair Sheet SRS67734

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Rennie Saverton

Inspiration Healthcare Ltd

Field Service Engineer

01455 840555

00012240

Rennie.Saverton@inspiration-healthcare.com

01/Apr/2021

Catherine Spence

Viamed

Warranty

Notes 16/Feb/2021 Steve Hardaker

16/Feb/2021 Steve Hardaker

16/Feb/2021 Steve Hardaker

Returning 3 more R-47V for investigation and replacement under warranty:

- IHC complaint ref CC994 - s/n V104589, supplied 28/10/2020

- IHC complaint ref CC996 s/n V104576, V104584, both supplied 28/10/2020

These need returning to Envitec urgently with a request for detailed failure report. Please see SN so that he can advise what is required.

01/Apr/2021 Catherine Spence

Received in

Ready For quote

Repair Complete Signed

CGNEEN 7.4.2021

replace ✓  
OK to go  
ahead.  
CG 6.4.2

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33509	Oxygen Sensor	0110047	V104576	Yes
SRN33510	Oxygen Sensor	0110047	V104584	Yes
SRN33511	Oxygen Sensor	0110047	V104589	Yes

111 days  
in  
service

0140013 x 3 @ £0

S/N, SRS, SRN

0110047 x 3 @ £0

new S/N, SRS, SRN

UPS x 1 @ £0.

please include

Ref:

CC994 - for V104589

CC996. for V104576/V10