

Service Repair Sheet SRS67726

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Chandra

Company/ Hospital Name

Royal Free Hospital

Department

Position

Direct Phone

020 78302417

General Phone

Opera Account

Email

Order Number

Date Received

09/Feb/2021

Booked in By

Catherine Spence

Main Company

Viamed

Type Return

Quote

Missed  
battery  
+  
calibration.  
Sent back to  
engineer.  
cg 22.2.21

Notes 09/Feb/2021 Catherine Spence

09/Feb/2021 Catherine Spence

Capnograph silicone cover S/N 81040152 dinger probe EFG01770

Broken screen needs repair and calibration also needs a new battery

Ready For quote

CGreen 22.2.21.

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33378	Capnograph	4410520	81040152	2
SRN33379	Pulse Oximetry Finger Sensor	0014750	EFG01770	2

4440000 x 0.25hr @ £90 per hr.  
S/N, SRS, SRN

4420590 x 1 £33.15  
SRS, SRN.

4430020 x 1 £9.90  
SRS, SRN

4480520 x 1 £150.  
SRS, SRN

\* 0040001 x 1 @ £0. NO fault found  
UPS x 1 £10. S/N, SRS, SRN

\* Please  
Add cg 7.4.21.