

Service Repair Sheet 67788

Contact Name	Ebaa Naji
Company/ Hospital Name	Inspiration Healthcare Ltd
Department	Quality Control
Position	Quality and Compliance Engineer
General Phone	01455 840555
Opera Account	00012240
Email	ebaa.naji@inspiration-healthcare.com
Date Received	08/Apr/2021
Booked in By	Catherine Spence

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33612	0110047	V104376	Oxygen Sensor	Complete - Repaired
				Time :0 Hour(s)

Parts Replaced

Qty: 1    0110047    Oxygen sensor R-47V

Replaced under warranty with S/N V105208

---