

Service Repair Sheet 67780

Contact Name	Laura Brewer
Company/ Hospital Name	J H M Butt and Co Ltd
Direct Phone	0130 271 8226
Opera Account	00020030
Email	laurab@buttsobawtry.com
Date Received	06/Apr/2021
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33517	0110132	556078	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0110132    Oxygen sensor - Automotive R-22AV

12.5 in air and 59.5 in oxygen but abnormal output. Replaced under warranty with S/N 577085

---