

Service Repair Sheet SRS67800

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Rennie Saverton

Inspiration Healthcare Ltd

Field Service Engineer

07825 941 145

01455 840555

00012240

Rennie.Saverton@inspiration-healthcare.com

19/Apr/2021

Robert Connor

Viamed

Warranty

Notes 09/Apr/2021 Steve Hardaker

09/Apr/2021 Steve Hardaker

09/Apr/2021 Steve Hardaker

Returning another R-47V, s/n V104579 from October 2020 batch. Outputs from 2 channels are mismatched and one channel is erratic.

Replace under warranty, return to Envitec.

09/Apr/2021 Steve Hardaker

09/Apr/2021 Steve Hardaker

09/Apr/2021 Steve Hardaker

09/Apr/2021 Steve Hardaker

Returning another R-47V, s/n V104579 from October 2020 batch. Outputs from 2 channels are mismatched and one channel is erratic.

Replace under warranty, return to Envitec.

19/Apr/2021 Robert Connor

1 x R47 s/n V104579

Ready For quote

Repair Complete Signed

CC Connor 19-4-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33618	Oxygen Sensor	0110047	V104579	Yes

0140013 x 1 @ EO
S/N, SRS, SEN0110047 x 1 @ ~~R47~~ EO.
New S/N, SRS, SEN

UPS x 1 @ EO

Add Customer
Ref CC1075